

ET99712 ET99711 (V) 微电脑 COD 多参数水质快速测定仪测量项目技术指标

代码	测量项目	测量范围	对应试剂	代码	测量项目	测量范围	对应试剂
20	需酸量	0.1 to 4.00 mmol/L	ET513210	223	总铁 TPTZ	0.02 to 1.80 mg/L Fe	ET530550
30	总碱度-M	5 to 200 mg/L CaCO ₃	ET513210	225	铁 LR	0.03 to 2.00 mg/L Fe	ET519990
31	总碱度-M HR	5 to 500 mg/L CaCO ₃	ET513240	226	铁 LR2	0.03 to 2.00 mg/L Fe	ET519492
35	总碱度-P	5 to 300 mg/L CaCO ₃	ET513230	227	铁 HR	0.1 to 10.0 mg/L Fe	ET519590
40	铝	0.01 to 0.30 mg/L Al	ET515460	240	锰	0.2 to 4.0 mg/L Mn	ET517621
			ET515470	242	锰 LR	0.01 to 0.70 mg/L Mn	ET535090
50	铝	0.01 to 0.25 mg/L Al	ET535000	注：水样中 CaCO ₃ 含量大于 300mg/L，选购 ET530640			
60	氨氮	0.02 to 1.00 mg/L N	ET512580	243	锰 HR	0.1 to 18.0 mg/L Mn	ET535100
			ET512590	245	锰	0.05 to 5.0 mg/L Mn	ET519055
注：海水水样，选购 ET460170 进行前处理，消除干扰				250	钼酸盐	1 to 50 mg/L MoO ₄	ET517631
62	氨氮 (pp)	0.01 to 0.80 mg/L N	ET535500	251	钼酸盐	0.05 to 5.00 mg/L MoO ₄	ET535450
65	氨氮 LR	0.02 to 2.50 mg/L N	ET535600	252	钼酸盐 HR	0.5 to 66.0 mg/L MoO ₄	ET535300
66	氨氮 HR	1 to 50 mg/L N	ET535650	254	钼酸盐 HR	1.0-100 mg/L MoO ₄	ET519790
70	双胍	2 to 60 mg/L PHMB	ET516100	257	镍	0.1-10 mg/L Ni	ET515630
80	溴	0.05 to 13.00 mg/L Br ₂	ET517711				ET515640
			注：高钙水样，选购 ET515740 取代 ET511050 试剂				ET465230
81	溴	0.05-4.5 mg/L Br ₂	ET530120	260	硝酸盐	0.08 to 1.00 mg/L N	ET502810
85	硼	0.1 to 2.0 mg/L B	ET517681				ET512310
90	氯化物	0.5 to 25.0 mg/L Cl	ET517741				注：需选购 ET366220 硝酸盐测试管
92	氯化物	0.5 to 20.0 mg/L Cl	ET519490	265	硝酸盐	1 to 30 mg/L N	ET535580
100	余氯	0.01 to 6.00 mg/L Cl ₂	ET511050	270	亚硝酸盐	0.01 to 0.50 mg/L N	ET512310
	总氯-亚氯酸盐	0.01 to 6.00 mg/L Cl ₂	ET517711	272	亚硝酸盐	0.01 to 0.30 mg/L N	ET530980
	注：测亚氯酸盐，需再订购 ET512170+ET512120+			280*	总氮 LR	0.5 to 25.0 mg/L N	ET535550
	ET511020；高钙水样，选购 ET515740 替换 ET511050			281*	总氮 HR	5 to 150 mg/L N	ET535560
101	余氯-总氯	0.02 to 4.00 mg/L Cl ₂	ET471056	290	活性氧	0.1 to 10.0 mg/L O ₂	ET511220
103	余氯-总氯	0.1 to 10.0 mg/L Cl ₂	ET511500	292	溶解氧	10 to 800 μg/L O ₂	ET380450
	注：测量总氯，使用 ET511500 试剂+ET511590 试剂			300	臭氧	0.02 to 2.00 mg/L O ₃	ET517711
105	高浓度氯	5 to 200 mg/L Cl ₂	ET517721		注：含氯水样，需选购 ET512170 试剂，消除干扰		
110	余氯	0.02 to 2.00 mg/L Cl ₂	ET530100	316	磷酸盐	0-125 mg/L PO ₄	ET535220
	总氯	0.02 to 2.00 mg/L Cl ₂	ET530120	320	正磷酸盐 LR	0.05 to 4.00 mg/L PO ₄	ET517651
111	余氯 HR	0.10 to 8.00 mg/L Cl ₂	ET530100	321	正磷酸盐 HR	1 to 80 mg/L PO ₄	ET517661
	总氯 HR	0.10 to 8.00 mg/L Cl ₂	ET530120	323	正磷酸盐 LR	0.06 to 2.50 mg/L PO ₄	ET531550
	注：测量需选购 ET197600 异形比色皿 (10mm 光程)			324	正磷酸盐 HR	0.06 to 5.00 mg/L PO ₄	ET535200
120	二氧化氯	0.02 to 11.00 mg/L ClO ₂	ET511050	325*	磷酸盐(酸解磷)	0.02 to 1.60 mg/L P	ET535250
	注：含氯水样，需选购 ET512170 试剂，消除干扰			326*	总磷-磷酸盐	0.02 to 1.10 mg/L P	ET535210

代码	测量项目	测量范围	对应试剂	代码	测量项目	测量范围	对应试剂
122	二氧化氯	0.04-3.80 mg/L ClO ₂	ET530100	327	磷酸盐 HR	5 to 40 mg/L PO ₄	ET380460
	注：含氯水样，需选购 ET512170 试剂，消除干扰			328	磷酸盐 LR	0.05 to 5.00 mg/L PO ₄	ET380480
125	六价铬	0.02 to 2.00 mg/L Cr ⁶⁺	ET537310	334	磷酸盐 LR	0.1-10 mg/L PO ₄	ET519765
	注：测量总铬*时，使用 ET537300+ET537310 试剂			335	磷酸盐	5 to 80 mg/L PO ₄	ET519090
130*	化学需氧量 LR	0 to 150 mg/L O ₂	ET99955	329	酸度 pH	5.2 to 6.8 pH	ET515700
131*	化学需氧量 MR	0 to 1500 mg/L O ₂	ET99974	330	酸度 pH	6.5 to 8.4 pH	ET511770
132*	化学需氧量 HR	0 to 15000 mg/L O ₂	ET99106	331	酸度 pH	6.5 to 8.4 pH	ET471040
150	余铜-总铜	0.05 to 5.0 mg/L Cu	ET517691	332	酸度 pH	8.0 to 9.6 pH	ET515710
151	余铜-总铜	0.05 to 4.0 mg/L Cu	ET519355	338	聚丙烯酸酯	1 to 30 mg/L Polyacryl	ET519165
153	余铜	0.05 to 5.00 mg/L Cu	ET530300	340	钾	0.7 to 12.0 mg/L K	ET515670
157	氰化物	0.01 to 0.50 mg/L CN	ET2418875	350	二氧化硅 LR	0.05 to 4.00 mg/L SiO ₂	ET517671
160	氰尿酸	0 to 160 mg/L CyA	ET511370		注：含磷酸盐水样，需选购 ET513150 试剂，消除干扰		
165	二乙基羟胺	0.02 to 0.5 mg/L DEHA	ET513220	351	二氧化硅 MR	0.1 to 1.6 mg/L SiO ₂	ET535690
			ET461185	352	二氧化硅 HR	1 to 90 mg/L SiO ₂	ET535700
167	二乙基羟胺	0.02 to 0.5 mg/L DEHA	ET536000	353	二氧化硅 HR	0.1-8.0 mg/L SiO ₂	ET519856
170	氟化物	0.05 to 2.00 mg/L F	ET467481	355	硫酸盐	5 to 100 mg/L SO ₄	ET515450
	注：首次须选购 ET205630 氟化物标准液，规格：30ml			360	硫酸盐	5 to 100 mg/L SO ₄	ET532160
190	钙硬度	50 to 900 mg/L CaCO ₃	ET515650	365	硫化物	0.04 to 0.50 mg/L S	ET502930
191	钙硬度	0 to 500 mg/L CaCO ₃	ET517761				ET502940
200	总硬度	2 to 50 mg/L CaCO ₃	ET515660	370	亚硫酸盐	0.1 to 5.0 mg/L SO ₃	ET518020
201	总硬度 HR	20 to 500 mg/L CaCO ₃	ET515660	380*	TOC LR	5.0 to 80.0mg/L TOC	ET420756
204	铂钴色度	0 to 500 mg/L Pt-Co	----	381*	TOC HR	50 to 800mg/L TOC	ET420756
205	联氨	0.05 to 0.50 mg/L N ₂ H ₄	ET462910	注：380 381 试剂首次使用，须选购 ET420757 专用消解帽			
	注：首次使用，须选购 ET384930 专用试剂勺			384	总固体悬浮物	0 to 750 mg/L TSS	----
206	联氨	0.005 to 0.60 mg/L N ₂ H ₄	ET531200	386	浊度	0 to 1000 FAU	----
207	联氨	0.01 to 0.70 mg/L N ₂ H ₄	ET380470	388	甲苯并三唑	1 to 16 mg/L Benzo-triazole	ET532200
210	过氧化氢	0.03 to 3.00 mg/L H ₂ O ₂	ET512380	390	尿素	0.1 to 2.5 mg/L Urea	ET517800
212	次氯酸钠	0.2-16% NaOCl	ET517721	400	锌	0.02 to 1.00 mg/L Zn	ET512620
213	过氧化氢	1-50 mg/L H ₂ O ₂	ET424991				ET512390
214	过氧化氢	10-500 mg/L H ₂ O ₂	ET424991	注：含高氯水样，需选购 ET512350 试剂，消除干扰			
215	碘	0.05 to 3.60 mg/L I ₂	ET511050	405	锌	0.1 to 2.5 mg/L Zn	ET519965
220	可溶性总铁	0.02 to 1.00 mg/L Fe	ET515370	900、910、920 mAbs(吸光度), 430、530、560nm			
	二价铁	0.02 to 1.00 mg/L Fe ²⁺	ET515420	930、940、950 mAbs(吸光度), 580、610、660nm			
222	铁	0.02 to 3.00 mg/L Fe	ET530560	*表示此项目需进行加热消解处理，请选购 ET99125(N)消解器			