

ET99732 ET99722(V) 微电脑 COD 多参数水质快速测定仪测量项目技术指标

代码	测量项目	测量范围	对应试剂	代码	测量项目	测量范围	对应试剂
20	需氧量	0.1 to 4.00 mmol/L	ET513210	222	铁	0.02 to 3.00 mg/L Fe	ET530560
30	总碱度-M	5 to 200 mg/L CaCO ₃	ET513210	223	总铁 TPTZ	0.02 to 1.80 mg/L Fe	ET530550
31	总碱度-M HR	5 to 500 mg/L CaCO ₃	ET513240	240	锰	0.2-4 mg/L Mn	ET517621
35	总碱度-P	5 to 300 mg/L CaCO ₃	ET513230	242	锰 LR	0.01 to 0.70 mg/L Mn	ET535090
40	铝	0.01 to 0.30 mg/L Al	ET517601		注：水样中 CaCO ₃ 含量大于 300mg/L，选购 ET530640		
50	铝	0.01 to 0.25 mg/L Al	ET535000	243	锰 HR	0.1 to 18.0 mg/L Mn	ET535100
60	氨氮	0.02 to 1.00 mg/L N	ET517611	250	钼酸盐	1 to 50 mg/L MoO ₄	ET517631
	注：海水水样，选购 ET460170 进行前处理，消除干扰			251	钼酸盐	0.05 to 5.00 mg/L MoO ₄	ET535450
62	氨氮 (pp)	0.01 to 0.80 mg/L N	ET535500	252	钼酸盐	0.5 to 66.0 mg/L MoO ₄	ET535300
63	氯胺	0.04 to 4.50 mg/L Cl ₂	ET531810	265	硝酸盐	1 to 30 mg/L N	ET535580
		0.01 to 0.50 mg/L NH ₄ -N		270	亚硝酸盐	0.01 to 0.50 mg/L N	ET512310
65	氨氮 LR	0.02 to 2.50 mg/L N	ET535600	272	亚硝酸盐	0.01 to 0.3 mg/L N	ET530980
66	氨氮 HR	1 to 50 mg/L N	ET535650	280*	总氮 LR	0.5 to 25.0 mg/L N	ET535550
70	双胍	2 to 60 mg/L PHMB	ET516100	281*	总氮 HR	5 to 150 mg/L N	ET535560
80	溴	0.05 to 13.00 mg/L Br ₂	ET511050	290	活性氧	0.1 to 10.0 mg/L O ₂	ET511220
	注：高钙水样，选购 ET515740 试剂取代 ET511050 试剂			292	溶解氧	10 to 800 μg/L O ₂	ET380450
85	硼	0.1 to 2.0 mg/L B	ET517681		首次使用需选购 ET192075 测量适配器		
90	氯化物	0.5 to 25.0 mg/L Cl ⁻	ET517741	300	臭氧	0.02 to 2.00 mg/L O ₃	ET517711
100	余氯-总氯	0.01 to 6.00 mg/L Cl ₂	ET511050		注：含氯水样，需选购 ET512170 试剂，消除干扰		
	注 1：测量总氯，使用 ET511050 试剂+ET511080 试剂			316	磷酸盐	0 to 125 mg/L PO ₄	ET535220
注 2：高钙水样，选购 ET515740 替换 ET511050			320	正磷酸盐 LR	0.05 to 4.00 mg/L PO ₄	ET517651	
101	余氯-总氯	0.02 to 4.00 mg/L Cl ₂	ET471056	321	正磷酸盐 HR	1 to 80 mg/L PO ₄	ET517661
103	余氯-总氯	0.1 to 10.0 mg/L Cl ₂	ET511500	323	正磷酸盐 LR	0.06 to 2.50 mg/L PO ₄	ET531550
	注：测量总氯，使用 ET511500 试剂+ET511590 试剂			324	正磷酸盐 HR	0.06 to 5.00 mg/L PO ₄	ET535200
105	高浓度氯	5 to 200 mg/L Cl ₂	ET517721	325*	磷酸盐(酸解磷)	0.02 to 1.60 mg/L P	ET535250
110	余氯	0.02 to 2.00 mg/L Cl ₂	ET530100	326*	总磷-磷酸盐	0.02 to 1.10 mg/L P	ET535210
	总氯	0.02 to 2.00 mg/L Cl ₂	ET530120	327	磷酸盐 HR	5 to 40 mg/L PO ₄	ET380460
120	二氧化氯	0.02 to 11.00 mg/L ClO ₂	ET511050	328	磷酸盐 LR	0.05 to 5.00 mg/L PO ₄	ET380480
	注：含氯水样，需选购 ET512170 试剂，消除干扰				首次使用需选购 ET192075 测量适配器		
130*	化学需氧量 LR	0 to 150 mg/L O ₂	ET99955	329	酸度 pH	5.2 to 6.8 pH	ET515700
131*	化学需氧量 MR	0 to 1500 mg/L O ₂	ET99974	330	酸度 pH	6.5 to 8.4 pH	ET511770
132*	化学需氧量 HR	0 to 15000 mg/L O ₂	ET99106	331	酸度 pH	6.5 to 8.4 pH	ET471040
150	余铜-总铜	0.05 to 5.00 mg/L Cu	ET517691	332	酸度 pH	8.0 to 9.6 pH	ET515710
153	余铜	0.05 to 5.00 mg/L Cu	ET530300	340	钾	0.7 to 12.0 mg/L K	ET515670

代码	测量项目	测量范围	对应试剂	代码	测量项目	测量范围	对应试剂
157	氰化物	0.01 to 0.50 mg/L CN	ET2418875	350	二氧化硅 LR	0.05 to 4.00 mg/L SiO ₂	ET517671
160	氰尿酸	0 to 160 mg/L CyA	ET511370		注：含磷酸盐水样，需选购 ET513150 试剂，消除干扰		
165	二乙基羟胺	20 to 500 µg/L DEHA	ET513220	351	二氧化硅 MR	0.1 to 1.6 mg/L SiO ₂	ET535690
			ET461185	352	二氧化硅 HR	1 to 90 mg/L SiO ₂	ET535700
167	二乙基羟胺	20 to 500 µg/L DEHA	ET536000	355	硫酸盐	5 to 100 mg/L SO ₄	ET515450
170	氟化物	0.05 to 2.00 mg/L F	ET467481	360	硫酸盐	5 to 100 mg/L SO ₄	ET532160
190	钙硬度	50 to 900 mg/L CaCO ₃	ET515650	365	硫化物	0.04 to 0.50 mg/L S	ET502930
191	钙硬度	0 to 500 mg/L CaCO ₃	ET517761				ET502940
200	总硬度	2 to 50 mg/L CaCO ₃	ET515660	370	亚硫酸盐	0.1 to 5.0 mg/L SO ₃	ET518020
201	总硬度 HR	20 to 500 mg/L CaCO ₃	ET515660	384	总固体悬浮物	0 to 750 mg/L TSS	----
204	铂钴色度	0 to 500 mg/L Pt	----	386	浊度	0 to 1000 FAU	----
205	联氨	0.05 to 0.50 mg/L N ₂ H ₄	ET462910	390	尿素	0.1 to 2.5 mg/L Urea	ET517611
	注：首次使用，须选购 ET384930 专用试剂勺						ET459300
206	联氨	0.005 to 0.60 mg/L N ₂ H ₄	ET531200				ET459400
207	联氨	0.01 to 0.70 mg/L N ₂ H ₄	ET380470	400	锌	0.02 to 1.00 mg/L Zn	ET512620
	首次使用需选购 ET192075 测量适配器						ET512390
210	过氧化氢	0.03 to 3.00 mg/L H ₂ O ₂	ET512380	注：含高氨水样，需选购 ET512350 试剂，消除干扰			
212	次氯酸钠	0.2 to 16% NaClO	ET517721	900、910、920 mAbs(吸光度)，430、530、560nm			
215	碘	0.05 to 3.60 mg/L I ₂	ET511050	930、940、950 mAbs(吸光度)，580、610、660nm			
220	可溶性总铁	0.02 to 1.00 mg/L Fe ³⁺	ET515370	* 表示此测量项目，需对样品进行加热消解处理，建议选购 ET99125(N)加热消解处理器			
	二价铁	0.02 to 1.00 mg/L Fe ²⁺	ET515420				

